

RESUSCITAIRE CHECKLIST – ALEX A&E

Top surface	
Heater Light Clock Clean, flat working surface O2 supply with pressure relief valve (replace if < 500 psi) Air supply with pressure relief valve (replace if <500 psi) Suction apparatus with Mini Yankeur	DAILY CHECK TO ENSURE WORKING
Towels	2
Hats – large & small	1 each
Stethoscope	1
Breathing Circuit with T-Piece	1
Self-inflating BVM bag (~500ml) with blow off valve	1 (hanging on side of resus)
O2 Saturation probe & sensor	1
Top Drawer	
Giraffe temperature skin probe(re-usable)	1
Giraffe heat reflective pads (disposable)	3
Cord clamp	1
Massimo adhesive sats sensor (disposable) Paediatric & Neonatal	2 of each
Masks – size 0,1	1 of each
Laryngoscopes Blades– size 00, 0, 1 (sterile).	1 of each
Laryngoscope handles (Green/metal)**DISPOSE OF AFTER USE (inc. batteries)	2
Endotracheal tube introducer	1
Endotracheal tubes – size 2.5, 3.0, 3.5	1 each
CO2 detector	1
Mini Yankeur/ Suction catheter – size 10	1 of each
Neobar ET Fixator – sizes Mini Ultra Micro Small Large X Large	1 of each size
Zinc Oxide Strappel tape	1
Sorbaderm Barrier Applicator	1
iGel size 1 with sachet of aquagel	1
Guedel Airway – size 1, 0, 00 & 000	1 of each
Bottom Drawer	
Box non-sterile gloves	1
Spare towels	2
Spare hats – large/ small	1 each
Spare breathing circuit with T-piece	1
Spare face masks size 0 and 1	1 each

Green O2 tubing	1
Plastic bag for pre-term infant	1
Emergency UVC Access	
Umbilical cord tie	1
Scalpel blade (disposable)	1
Spencer wells forceps (7ins)	1
Umbilical venous catheters – sizes 3.5 & 5	2 of each
3-way tap	1
Silk suture 3-0	1
Adhesive Tape for umbilical lines (Elastoplast)	1 roll
Other	
Trans warmer	1
Maternity Delivery pack	2
New-born Life Support Algorithm (in folder)	1
Signature checking list/ spares (in folder)	

Emergency Drug Box Kept in Resus Cupboard